

MCScap LIGHT textile cap

Printed textile cap marked according to international system 10-10 with fixing rubber.

REF	Size	Head circumference
003-0-284	XL	60-66 cm
003-0-285	XL/L	57-63 cm
003-0-286	L	54-60 cm
003-0-287	L/M	51-57 cm
003-0-288	M	48-54 cm
003-0-289	M/S	45-51 cm
003-0-290	S	42-48 cm
003-0-291	S/XS	39-45 cm
003-0-292	XS	36-42 cm
003-0-293	Inf I	32-36 cm
003-0-294	Inf II	28-32 cm
003-0-295	Inf III	24-28 cm



SET

- MCScap LIGHT textile cap.
- operational documentation.

DESCRIPTION

MCScap LIGHT is a textile cap marked according to the international 10-10 system. The cap is designed for use with MCScap-C/TC/T/NT/NTC/NTH/StTC/StNT electrodes. Textile cap is made of elastic material, preserving the shape and size. The cap provides the exact position of the electrodes on the head without additional measurements and adjustments. Large holes are provided for ventilation and access to the electrodes and patient's skin. The cap is fixed on the head with an elastic band. Size identification is carried out by the color of the seam.

SPECIFICATION

Material of the cap	neoprene
Electrode positioning system	10-10
Compatible Electrode Models	MCScap-C/TC/T/NT/NTC/NTH/StTC/StNT
Color	black
Net weight	< 30 g
Gross weight	< 30 g



MCSCAP SIZE

Cap size	Head circumference	Color of cap for colored caps	Color of seams for black caps	Intended* for	Size of chest or chin belt
XL	60-66 cm	green 	green 	adults	adults
XL/L	57-63 cm	green and blue 	green and blue 	adults	
L	54-60 cm	blue 	blue 	adults(most)	
L/M	51-57 cm	blue and red 	blue and red 	teenagers, adults	
M	48-54 cm	red 	red 	children, teenagers	children
M/S	45-51 cm	red and yellow 	red and yellow 	children up to 5 years	
S	42-48 cm	желтый 	yellow 	children up to 2 years	
S/XS	39-45 cm	yellow and green 	yellow and green 	children up to 1 year	infant
XS	36-42 cm	green 	green 	babies up to 4 month	
Inf I	32-36 cm	blue 	blue 	infants up to 1 month	
Inf II	28-32 cm	red 	red 	newborn infants, preemies	
Inf III	24-28 cm	желтый 	yellow 	newborn infants, preemies	

* Recommendations are based on WHO anthropological data <http://www.who.int/childgrowth> and ISO/TR 7250-2:2010. This information is a reference. In each case, the size of the cap should be selected, guided by the measured value of the circumference of the patient's head.



Before examination, measure the patient's head circumference and select a cap of the appropriate size. Choosing the correct cap size is important for correct EEG recording and patient comfort.

