

## MCScap COVER reusable protective cover

Protective cap for MCScap® EEG caps.

REF	Size
003-0-272	XL
003-0-273	XL/L
003-0-274	L
003-0-275	L/M
003-0-276	M
003-0-277	M/S
003-0-278	S
003-0-279	S/XS
003-0-280	XS



### SET

- MCScap COVER,
- operational documentation.

### DESCRIPTION

MCScap COVER is designed for:

- additional fixation of the EEG cap,
- better fit of the electrodes to the head,
- protection of wires with electrodes from external factors.

MCScap COVER is suitable for both caps with location of the output of the electrode cable from cap on the back of the head (CLINIC, PROFESSIONAL, ELECTRODE SET), and at the top (SLEEP).

The size range of the MCScap COVER is identical to the size range of MCScap® textile caps. MCScap COVER is recommended to use the same size as the cap. For ELECTRODE SET kits it is recommended to use a half-size larger cap.

The MCScap COVER is worn over the top of the electrode cap, attached to the cap velcro and secured with chin or chest belt. The chin belt is not included in the MCScap COVER.

### SPECIFICATION

Material of the cap	compression net material
Color	black
Net weight	< 18 g
Gross weight	< 18 g



## MCSCAP SIZE

Cap size	Head circumference	Color of cap for colored caps	Color of seams for black caps	Intended* for	Size of chest or chin belt
XL	60-66 cm	green 	 green	adults	adults
XL/L	57-63 cm	green and blue 	 green and blue	adults	
L	54-60 cm	blue 	 blue	adults(most)	
L/M	51-57 cm	blue and red 	 blue and red	teenagers, adults	
M	48-54 cm	red 	 red	children, teenagers	children
M/S	45-51 cm	red and yellow 	 red and yellow	children up to 5 years	
S	42-48 cm	желтый 	 yellow	children up to 2 years	
S/XS	39-45 cm	yellow and green 	 yellow and green	children up to 1 year	infant
XS	36-42 cm	green 	 green	babies up to 4 month	
Inf I	32-36 cm	blue 	 blue	infants up to 1 month	
Inf II	28-32 cm	red 	 red	newborn infants, preemies	
Inf III	24-28 cm	желтый 	 yellow	newborn infants, preemies	

\* Recommendations are based on WHO anthropological data <http://www.who.int/childgrowth> and ISO/TR 7250-2:2010. This information is a reference. In each case, the size of the cap should be selected, guided by the measured value of the circumference of the patient's head.



*Before examination, measure the patient's head circumference and select a cap of the appropriate size. Choosing the correct cap size is important for correct EEG recording and patient comfort.*

