

DATA SHEET

MCScap 10-5 textile cap

Printed textile cap marked according to international system 10-5.

REF	Size	Head circumference
003-0-264	L	54-60 cm
003-0-265	L/M	51-57 cm
003-0-354	M	48-54 cm



SET

- MCScap 10-5 textile cap,
- chin belt,
- operational documentation.

DESCRIPTION

MCScap 10-5 is a textile cap marked according to the international 10-5 system. The cap is designed for use with MCScap-C/TC/T/NT/NTC/NTH/StTC/StNT electrodes. Textile cap is made of elastic material, preserving the shape and size. The cap provides the exact position of the electrodes on the head without additional measurements and adjustments. Large holes are provided for ventilation and access to the electrodes and patient's skin. The cap is fixed on the head with the chin or chest belt. Size identification is carried out by the color of the seam.

SPECIFICATION

Material of the cap	neoprene
Electrode positioning system	10-5
Compatible Electrode Models	MCScap-C/TC/T/NT/NTC/NTH/StTC/StNT
Color	black
Net weight	<40 g
Gross weight	< 40 g

MCSCAP SIZE

Cap size	Head circumference	Color of cap for colored caps		Color	of seams for black caps	Intended* for	Size of chest or chin belt
XL	60-66 cm	green			green	adults	adults
XL/L	57-63 cm	green and blue			green and blue	adults	
L	54-60 cm	blue			blue	adults(most)	
L/M	51-57 cm	blue and red			blue and red	teenagers, adults	
М	48-54 cm	red			red	children, teenagers	children
M/S	45-51 cm	red and yellow			red and yellow	children up to 5 years	
S	42-48 cm	желтый			yellow	children up to 2 years	
S/XS	39-45 cm	yellow and green			yellow and green	children up to 1 year	
XS	36-42 cm	green			green	babies up to 4 month	infant
Infl	32-36 cm	blue			blue	infants up to 1 month	
InfII	28-32 cm	red			red	newborn infants, preemies	
InfIII	24-28 cm	желтый			yellow	newborn infants, preemies	

^{*} Recommendations are based on WHO anthropological data http://www.who.int/childgrowth and ISO/TR 7250-2:2010. This information is a reference. In each case, the size of the cap should be selected, guided by the measured value of the circumference of the patient's head.

Before examination, measure the patient's head circumference and select a cap of the appropriate size. Choosing the correct cap size is important for correct EEG recording and patient comfort.